

CERTIFICATE of CONFORMATY

MR:comp GmbH herewith declares that the MR safety & MR compatibility of the product listed below has been evaluated according to the applicable standard practices IEC 62570 / ASTM F2503 for medical devices within a magnet resonance environment.

The product has been classified as
MR conditional for use within a magnet resonance environment up to 3.0 Tesla.

Product **Titanium tools (100 % titanium 6AL-4V)**

Contact **Lucas Mason**
General Manager & Executive Vice President

Manufacturer




Imprex International, Inc.,
2916 Malmo Road,
60005 Arlington Heights, IL
United States


Classification
IEC 62570
ASTM F2503



MR Conditional

Test Report no. **TR0384-101a (ASTM F2503)**
TR0384-301 (ASTM F2052)


Wolfgang Görtz, Dipl.-Phys.
Head of Test Laboratory


Dr. rer. nat. Nirav Darji
Head of Consulting Department